



Government of India
Bhabha Atomic Research Centre
Nuclear Recycle Group
Process Development Division

Ref: PSDD/MDS/Lab/2017/02A/130008

Date: 24/07/2017

26

Sub: Maintenance of existing two fume hoods, cabin doors and painting of existing cupboard & cabinets as per our technical specification.

Dear,

You are requested to submit your quotation in sealed envelope for the above mentioned job. The material should confirm to our specifications. The reference no. given above should be clearly mentioned on the sealed envelope.

Quotation shall be complete in all respects with regard to specifications, validity of offer etc., and must reach the following address on or before **11/08/2017** by 14:00 hr.

S. A. KHOT
SA/E, PSDD,
Room no. 211, PSDD (LAB), WIP Complex,
BARC, Trombay, Mumbai-400 085
Tel: 022-25596727

Due consideration shall be given to the following aspects while you submit your offer:

1. The work shall be completed within 30 days from issue of work order.
2. The offer shall be valid for a period of thirty days and in case of placement of the work order, shall remain firm till the completion of the work.
3. Quotation is to be in printed letter head/quotation format which should consist of Sales Tax Registration Number registered with local ST authority/ CST authority, PAN Number of the firm, Service Tax Registration Number, etc.
4. Quotation shall be sent preferably by **speed post only**.
5. It shall be noted that BARC is exempted for ED and Octroi.
6. Taxes, duties and other charges applicable, if any, shall be indicated separately.
7. Contactors are requested to provide complete details of materials to be supplied with make.
8. The contractor shall arrange all necessary equipments and personnel required for the successful completion of the work and the department will not release any free-issue material except electricity and water.
9. Contractors are advised to visit the site of work, before quoting rates for the items to acquaint themselves with site conditions.
10. Work will be carried out at Room no.211, WIP Process Building, BARC, Trombay.


S. A. KHOT
SA/E, PSDD

Technical Specification

Maintenance of existing two fume hoods, cabin doors and painting of existing cupboard & cabinets as per our technical specification.

1. Maintenance of fume hoods: 2No.

Maintenance of two fume hoods, which includes operational of doors & replacement of existing Perspex sheets, lightening in fume hoods, old electrical switches and plugs , replacement of connections for air & water inlet systems with new one, painting from inside with epoxy paint ,outside fume hood and stand with oil paint.

Sr.	Item	Description
1	Perspex sheet for fume hoods	Size:1000mmX830mmX5mm
2	Interior Paint for fume hood	Epoxy paint only
3	Electrical switches and plugs	Anchor or other with ISI mark
4	Connection for air and water in fume hood; 4no.	SS tubing with valve


2. Maintenance of Cabin doors: 3No.

Repairing of three existing aluminium cabin doors including joint fixing with rod or wood filling inside sections, hinges, locking arrangement and door closers.

3. Spray painting of steel cupboard/cabinets and electrical pannel

Spray painting of existing cupboard and cabinets, electrical panel, racks etc. in PSDD lab. Room No.211 with one coat of red oxide and two coats of oil paint. Approx. area to be painted 250 sq.m.

4. No Free Issue Material will be given by BARC except Electricity and Water.


S. A. Khot
SA/E, PSDD
24/07/2017